

System type:	LC00190-A
Serial number:	000045
Date:	27.08.2026

INSPECTION POINTS	Accept	Reject	Note
System on/off, system start	☒	☐	
Control check	☒	☐	
Rotation speed stability test			
500 rpm	☒	☐	
1000 rpm	☒	☐	
1500 rpm	☒	☐	
2000 rpm	☒	☐	
Standing rotor test			
50 bar	☒	☐	
75 bar	☒	☐	
100 bar	☒	☐	
Water flow rate test			
4 ml/min - measured pressure: bar 1	☒	☐	
8 ml/min - measured pressure: bar 1	☒	☐	
12 ml/min - measured pressure: bar 1	☒	☐	
Measured rotor volume: ml	135 ☒	☐	
Test chromatogram	☒	☐	

	Prepared By	Approved By
Name	TAMÁS RÖZSÁVÖLCEI	
Signature		
Date	27.08.2026.	