



System type:	LC00190-A
Serial number:	000041
Date:	27. 04. 2026.

INSPECTION POINTS	Accept	Reject	Note
System on/off, system start	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Control check	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Rotation speed stability test			
500 rpm	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
1000 rpm	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
1500 rpm	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
2000 rpm	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Standing rotor test			
50 bar	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
75 bar	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
100 bar	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Water flow rate test			
4 ml/min - measured pressure: bar	1 <input checked="" type="checkbox"/>	<input type="checkbox"/>	
8 ml/min - measured pressure: bar	1 <input checked="" type="checkbox"/>	<input type="checkbox"/>	
12 ml/min - measured pressure: bar	2 <input checked="" type="checkbox"/>	<input type="checkbox"/>	
Measured rotor volume: ml	135 <input checked="" type="checkbox"/>	<input type="checkbox"/>	
Test chromatogram	<input checked="" type="checkbox"/>	<input type="checkbox"/>	

	Prepared By	Approved By
Name	TAMÁS RÓZSAVÖLGTI	AHILKA DECSI
Signature	<i>[Handwritten Signature]</i>	<i>[Handwritten Signature]</i>
Date	27. 04. 2026.	27. 04. 2026