



System type:	LC00090-F
Serial number:	000046
Date:	21.09.2026.

INSPECTION POINTS	Accept	Reject	Note
System on/off, system start	☒	☐	
Control check	☒	☐	
Rotation speed stability test			
500 rpm	☒	☐	
1000 rpm	☒	☐	
1500 rpm	☒	☐	
2000 rpm	☒	☐	
Standing rotor test			
50 bar	☒	☐	
75 bar	☒	☐	
100 bar	☒	☐	
Water flow rate test			
1 ml/min - measured pressure: bar	1 ☒	☐	
2 ml/min - measured pressure: bar	1 ☒	☐	
3 ml/min - measured pressure: bar	2 ☒	☐	
Measured rotor volume: ml	35 ☒	☐	
Test chromatogram	☒	☐	

	Prepared By	Approved By
Name	TAMÁS RÓZSAVÖLCSEI	Orosz Máté
Signature		
Date	21.09.2026.	24.09.2026.