



System type:	LC00090-F
Serial number:	000034
Date:	02.04.2026.

INSPECTION POINTS	Accept	Reject	Note
System on/off, system start	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Control check	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<b>Rotation speed stability test</b>			
500 rpm	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
1000 rpm	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
1500 rpm	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
2000 rpm	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<b>Standing rotor test</b>			
50 bar	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
75 bar	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
100 bar	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<b>Water flow rate test</b>			
1 ml/min - measured pressure: bar	1 <input checked="" type="checkbox"/>	<input type="checkbox"/>	
2 ml/min - measured pressure: bar	1 <input checked="" type="checkbox"/>	<input type="checkbox"/>	
3 ml/min - measured pressure: bar	2 <input checked="" type="checkbox"/>	<input type="checkbox"/>	
Measured rotor volume: ml	34 <input checked="" type="checkbox"/>	<input type="checkbox"/>	
Test chromatogram	<input checked="" type="checkbox"/>	<input type="checkbox"/>	

	Prepared By	Approved By
Name	TAMÁS RÓZSAVÖLÁGI	ATILKA DECSI
Signature		
Date	02.04.2026.	02.04.2026