



System type:	LC000090-F
Serial number:	000037
Date:	2026.04.01

INSPECTION POINTS	Accept	Reject	Note
System on/off, system start	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	
Control check	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	
Rotation speed stability test			
500 rpm	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	
1000 rpm	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	
1500 rpm	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	
2000 rpm	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	
Standing rotor test			
50 bar	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	
75 bar	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	
100 bar	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	
Water flow rate test			
1 ml/min - measured pressure: 0 bar	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	0 bar
2 ml/min - measured pressure: 0 bar	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	0 bar
3 ml/min - measured pressure: 1 bar	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	1 bar
Measured rotor volume: ml 32	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	32 ml
Test chromatogram	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	

	Prepared By	Approved By
Name	TAMÁS RÓZSAVÖLGYI	ATTILA DECSI
Signature		
Date	2026.04.01.	2026.04.01