



LiLiCHRO

miniLiLi LLC Column V5.0

Factory Acceptance Test

Version: 5.0
Author: László Németh
Date: 2024.03.01.
P/N: LC00090-F

Disclosure

This document is confidential and is the property of LiLiChro Ltd.
No attempt should be made to copy this document in any way without prior consent.
©2022 LiLiChro - All rights reserved



System type:	LC00090-F
Serial number:	000042
Date:	24.06.2026.

INSPECTION POINTS	Accept	Reject	Note
System on/off, system start	☒	☐	
Control check	☒	☐	
Rotation speed stability test			
500 rpm	☒	☐	
1000 rpm	☒	☐	
1500 rpm	☒	☐	
2000 rpm	☒	☐	
Standing rotor test			
50 bar	☒	☐	
75 bar	☒	☐	
100 bar	☒	☐	
Water flow rate test			
1 ml/min - measured pressure: bar	1 ☒	☐	
2 ml/min - measured pressure: bar	1 ☒	☐	
3 ml/min - measured pressure: bar	2 ☒	☐	
Measured rotor volume: ml	34 ☒	☐	
Test chromatogram	☒	☐	

	Prepared By	Approved By
Name	TAMÁS RÓZSAVÖLGYI	Attila DECSI
Signature	Uti T.	Attila DECSI
Date	24.06.2026.	24.06.2026.

Test chromatogramm Mini000042

